

FINEVISION TRIUMF

EDOF TRIFOCAL OPTIC



PhysIOL


ADVANCED OPTICAL SOLUTIONS



FINEVISION TRIUMF

EDOF trifocal
hydrophobic IOL

Technical specifications

Commercial name	Pod L GF		
Material	PhysIOL G-free® (hydrophobic acrylic glistening-free)*		
LCA	Chromatic aberration-free** 		
Overall diameter	11.40 mm		
Optic diameter	6.00 mm		
Optic	Biconvex aspheric (-0.11μ SA)		
Haptic design	Double C-loop & RidgeTech®		
Filtration	UV & blue light		
Refractive index	1.52		
Abbe number	42		
Angulation	5°		
Additional power	Elongated depth of focus energy with +1.75D & +3.50D addition		
Injection system	Medicel Accuject 2.0 from 10D to 24.5D Medicel Accuject 2.1 / 2.2 from 25D to 35D		
Incision size	≥ 2.0 mm		
Spherical power	10D to 35D (0.5D steps)		
Square edge	360°		
Nominal manufacturer A constant	119.40		
Suggested A constant***		Interferometry	Ultrasound
	Hoffer Q: pACD	5.85	5.59
	Holladay 1: Sf	2.06	1.80
	Barrett: LF	2.09	-
	SRK/T: A	119.40	119.05
	Haigis****: a0; a1; a2	1.70; 0.4; 0.1	1.214; 0.4; 0.1

* The PhysIOL G-free® is patented since 2010.

** For far and intermediate vision.

*** Estimates only: surgeons are recommended to use their own values based upon their personal experience. Refer to our website for updates.

**** Not optimized.

INJECTION GUIDELINES







The Medical Accuject injection system is recommended for implanting the FineVision^{Triumf} lenses.

This fully single-use system represents total reliability for safe and effective lens injections.

Its compact design with integrated cartridge enables a simple, predictable loading and positioning of the lens.


Accuject 2.0 for lens diopters < 25D

Accuject 2.1 or 2.2 for lens diopters ≥ 25D



1. Apply viscoelastic into the tip and the loading chamber of the injector cartridge.
2. Remove the lens from the lens holder. Position the lens into the cartridge in such way that the two haptics with the holes are pointing at 1 and 7 o'clock.
3. Exert slight pressure onto the lens optic and make sure that all haptics are inside before further closing the cartridge. Close the cartridge and check the position of the lens.
4. Once the "click-lock" mechanism engages, the lens is securely loaded and ready for injection.
5. Press the injector plunger forward and push the lens into the conical tip of the cartridge.
6. Pull the plunger back a few millimeters and then inject the lens in one continuous motion. For gently implantation, it is not necessary to push the plunger until the end of the cartridge.

RIDGETECH



The **RidgeTech**[®] design reduces the risk of stickiness between the haptics and the optic.

It ensures a safe injection and reliable unfolding of the lens.

Distributed by



PhysIOL sa/nv - Liège Science Park - Allée des Noisetiers 4 - 4031 Liège - Belgium
t. +32 (0)4 361 05 49 - f. +32 (0)4 361 05 30 - info@physiol.be - www.physiol.eu